

The North Carolina Polygraph Association



“Dedicated to the improvement of the qualifications of Polygraph Examiners through high standards of professional ethics, conduct, education, and achievement.”

Amanda Nosalek, President
Kevin Gregson, Vice-President
John Vanderbilt, Secretary/Treasurer

Andrew Goldstein, Private Sector Representative
Andrew Newsom, Law Enforcement Representative
Frank Suddreth, Training Coordinator
Dustin Nowatka, Board Chairman

The NCPA website is located at: <http://northcarolinapolygraphassociation.org/>

Dear NCPA members,

We have received questions from members looking for guidance on conducting polygraph examinations during the current COVID-19 pandemic.

The president, our governor and experts have said we should all be practicing social distancing which they have defined as being at least six feet from others. Due to the proximity between an examiner and examinee this cannot be accomplished when conducting a polygraph examination.

We recommend not conducting any polygraph examinations unless necessary until social distancing is no longer the recommendation.

If you do have a critical polygraph examination that needs to be conducted here are some guidelines for how to clean your equipment from Lafayette.

The Lafayette LX6, LX5000, and LX4000 systems and all accessories may be cleaned with any the following:

- Alcohol wipes*
- Disinfectant wipes*
- Damp towel with anti-bacterial soap*

Cautions:

- Do not use bleach on any part of the polygraph system.*
- Wipe down the surfaces of the equipment only, do not soak or saturate any part of the system with liquid.*
- Do not use alcohol, disinfectant (or disinfectant wipes), or bleach on the rubber O-rings of the cardio or pneumo connectors.*

If you have any questions, please do not hesitate to reach out to any of the NCPA board members.

Take care of yourselves and your families.

Sincerely,

NCPA President Amanda Nosalek